



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Nationwide Gage Calibration, Inc.

**135 East Saint Charles Road, Suite E
Carol Stream, IL 60188**

Fulfills the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 28 March 2027

Certificate Number: AC-1160



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Nationwide Gage Calibration, Inc.

135 East Saint Charles Road, Suite E

Carol Stream, IL 60188

Richard Parish 630-529-4959

CALIBRATION

Valid to: **March 28, 2027**

Certificate Number: **AC-1160**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source	Up to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 020) V	7.8 μ V 43 μ V 0.47 mV 6.1 mV 22 mV	Comparison to Multi-Product Calibrator
DC Current – Source	Up to 330 μ A (0.33 to 3.3) mA (3.3 to 33) mA (33 to 330) mA (0.33 to 1.1) A (1.1 to 3) A (3 to 11) A (11 to 20.5) A	0.1 μ A 0.39 μ A 3.9 μ A 39 μ A 0.24 mA 1.4 mA 6.4 mA 24 mA	Comparison to Fluke 5520A Multi-Product Calibrator
Resistance – Source	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) k Ω (1.1 to 3.3) k Ω (3.3 to 11) k Ω (11 to 33) k Ω (33 to 110) k Ω (110 to 330) k Ω (0.33 to 1.1) M Ω (1.1 to 3.3) M Ω	0.97 m Ω 1.5 m Ω 3.7 m Ω 14 m Ω 37 m Ω 0.11 Ω 0.37 Ω 1.1 Ω 3.7 Ω 13 Ω 38 Ω 0.24 k Ω	Comparison to Multi-Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source	(3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ (0.33 to 1.1) GΩ	1.7 kΩ 9.8 kΩ 65 kΩ 1.2 MΩ 2.4 MΩ	Comparison to Multi-Product Calibrator
AC Voltage – Source	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (0.33 to 3.3) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (33 to 330) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (330 to 1 020) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	31 μV 6.6 μV 8.3 μV 39 μV 0.14 mV 0.31 mV 0.13 mV 57 μV 62 μV 0.14 mV 0.31 mV 0.77 mV 1.2 mV 0.58 mV 0.73 mV 1.2 mV 2.7 mV 9.2 mV 12 mV 5.8 mV 9.2 mV 14 mV 35 mV 74 mV 78 mV 97 mV 0.12 V 0.77 V 0.35 V 0.3 V 0.36 V	Comparison to Multi-Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source	(29 to 330) μ A		Comparison to Multi-Product Calibrator
	(10 to 20) Hz	1.2 μ A	
	(20 to 45) Hz	0.99 μ A	
	45 Hz to 1 kHz	0.94 μ A	
	(1 to 5) kHz	1.4 μ A	
	(5 to 10) kHz	3.2 μ A	
	(10 to 30) kHz	6.2 μ A	
	(0.33 to 3.3) mA		
	(10 to 20) Hz	7.8 μ A	
	(20 to 45) Hz	4.9 μ A	
	45 Hz to 1 kHz	4 μ A	
	(1 to 5) kHz	7.7 μ A	
	(5 to 10) kHz	20 μ A	
	(10 to 30) kHz	39 μ A	
	(3.3 to 33) mA		
	(10 to 20) Hz	70 μ A	
	(20 to 45) Hz	36 μ A	
	45 Hz to 1 kHz	18 μ A	
	(1 to 5) kHz	32 μ A	
	(5 to 10) kHz	78 μ A	
	(10 to 30) kHz	0.16 mA	
	(33 to 330) mA		
	(10 to 20) Hz	0.69 mA	
	(20 to 45) Hz	0.35 mA	
	45 Hz to 1 kHz	0.16 mA	
	(1 to 5) kHz	0.39 mA	
	(5 to 10) kHz	0.77 mA	
	(10 to 30) kHz	1.6 mA	
	(0.33 to 1.1) A		
	(10 to 45) Hz	2.1 mA	
	45 Hz to 1 kHz	0.59 mA	
	(1 to 5) kHz	7 mA	
	(5 to 10) kHz	29 mA	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source	(1.1 to 3) A		Comparison to Multi-Product Calibrator
	(10 to 45) Hz	6.3 mA	
	45 Hz to 1 kHz	2.3 mA	
	(1 to 5) kHz	21 mA	
	(5 to 10) kHz	87 mA	
	(3 to 11) A		
	(45 to 100) Hz	8.1 mA	
	100 Hz to 1 kHz	13 mA	
	(1 to 5) kHz	0.39 A	
	(11 to 20.5) A		
Capacitance – Source	(45 to 100) Hz	29 mA	Comparison to Multi-Product Calibrator
	100 Hz to 1 kHz	35 mA	
	(1 to 5) kHz	0.7 A	
	(190 to 400) pF	6.3 pF	
	(0.4 to 1.1) nF	6.5 pF	
	(1.1 to 3.3) nF	20 pF	
	(3.3 to 11) nF	34 pF	
	(11 to 33) nF	0.13 nF	
	(33 to 110) nF	0.34 nF	
	(110 to 330) nF	1.3 nF	
	(0.33 to 1.1) μ F	3.4 nF	
	(1.1 to 3.3) μ F	13 nF	
	(3.3 to 11) μ F	34 nF	
	(11 to 33) μ F	0.21 μ F	
	(33 to 110) μ F	0.61 μ F	
	(110 to 330) μ F	2.3 μ F	
	(0.33 to 1.1) mF	6.1 μ F	
	(1.1 to 3.3) mF	18 μ F	
Electrical Simulation of Thermocouple Indicating Devices	(3.3 to 11) mF	60 μ F	Comparison to Multi-Product Calibrator
	(11 to 33) mF	0.19 mF	
	(33 to 110) mF	0.97 mF	
	Type B		
	(600 to 800) $^{\circ}$ C	0.46 $^{\circ}$ C	
	(800 to 1 000) $^{\circ}$ C	0.37 $^{\circ}$ C	
	(1 000 to 1 550) $^{\circ}$ C	0.33 $^{\circ}$ C	
	(1 550 to 1 820) $^{\circ}$ C	0.36 $^{\circ}$ C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices	Type C		Comparison to Multi-Product Calibrator
	(0 to 150) °C	0.33 °C	
	(150 to 650) °C	0.29 °C	
	(650 to 1 000) °C	0.34 °C	
	(1 000 to 1 800) °C	0.52 °C	
	(1 800 to 2 316) °C	0.85 °C	
	Type E		
	(-250 to -100) °C	0.52 °C	
	(-100 to -25) °C	0.21 °C	
	(-25 to 350) °C	0.19 °C	
	(350 to 650) °C	0.21 °C	
	(650 to 1 000) °C	0.25 °C	
	Type J		
	(-210 to -100) °C	0.3 °C	
	(-100 to -30) °C	0.21 °C	
	(-30 to 150) °C	0.19 °C	
	(150 to 760) °C	0.21 °C	
	(760 to 1 200) °C	0.27 °C	
	Type K		
	(-200 to -100) °C	0.36 °C	
	(-100 to -25) °C	0.22 °C	
	(-25 to 120) °C	0.21 °C	
	(120 to 1 000) °C	0.29 °C	
	(1 000 to 1 372) °C	0.42 °C	
	Type L		
	(-200 to -100) °C	0.39 °C	
	(-100 to 800) °C	0.29 °C	
	(800 to 900) °C	0.21 °C	
	Type N		
	(-200 to -100) °C	0.42 °C	
	(-100 to -25) °C	0.26 °C	
	(-25 to 120) °C	0.23 °C	
	(120 to 410) °C	0.22 °C	
	(410 to 1 300) °C	0.3 °C	
	Type R		
	(0 to 250) °C	0.59 °C	
	(250 to 400) °C	0.38 °C	
	(400 to 1 000) °C	0.36 °C	
	(1 000 to 1 767) °C	0.42 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices	Type S (0 to 250) °C	0.49 °C	Comparison to Multi-Product Calibrator
	(250 to 1 000) °C	0.38 °C	
	(1 000 to 1 400) °C	0.39 °C	
	(1 400 to 1 767) °C	0.48 °C	
	Type T (-250 to -150) °C	0.65 °C	
	(-150 to 0) °C	0.27 °C	
	(0 to 120) °C	0.21 °C	
	(120 to 400) °C	0.19 °C	
	Type U (-200 to 0) °C	0.58 °C	
	(0 to 600) °C	0.3 °C	
Electrical Simulation of Thermocouple Indicators ¹	Type B (600 to 1 800) °C	2.6 °C	Comparison to Precision Temperature Calibrator
	Type E (-200 to 950) °C	1.1 °C	
	Type J (-200 to 1 200) °C	1.3 °C	
	Type K (-200 to 1 370) °C	1.5 °C	
	Type L (-200 to 1 370) °C	1.1 °C	
	Type N (-200 to 1 370) °C	1.8 °C	
	Type R (-20 to 1 750) °C	3 °C	
	Type S (-20 to 1 750) °C	3 °C	
	Type T (-200 to 400) °C	1.5 °C	
	Type U (-200 to 400) °C	1.4 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices	Pt 385, 100 Ω		Comparison to Multi-Product Calibrator
	(-200 to -80) °C	0.12 °C	
	(-80 to 0) °C	0.12 °C	
	(0 to 100) °C	0.13 °C	
	(100 to 300) °C	0.14 °C	
	(300 to 400) °C	0.15 °C	
	(400 to 630) °C	0.16 °C	
	(630 to 800) °C	0.26 °C	
	Pt 3926, 100 Ω		
	(-200 to -80) °C	0.12 °C	
	(-80 to 0) °C	0.12 °C	
	(0 to 100) °C	0.13 °C	
	(100 to 300) °C	0.14 °C	
	(300 to 400) °C	0.15 °C	
	(400 to 630) °C	0.16 °C	
	Pt 3916, 100 Ω		
	(-200 to -190) °C	0.27 °C	
	(-190 to -80) °C	0.11 °C	
	(-80 to 0) °C	0.12 °C	
	(0 to 100) °C	0.12 °C	
	(100 to 260) °C	0.13 °C	
	(260 to 300) °C	0.13 °C	
	(300 to 400) °C	0.14 °C	
	(400 to 600) °C	0.15 °C	
	(600 to 630) °C	0.26 °C	
	Pt 385, 200 Ω		
	(-200 to -80) °C	0.11 °C	
	(-80 to 0) °C	0.11 °C	
	(0 to 100) °C	0.11 °C	
	(100 to 260) °C	0.12 °C	
	(260 to 300) °C	0.16 °C	
	(300 to 400) °C	0.17 °C	
	(400 to 600) °C	0.18 °C	
	(600 to 630) °C	0.19 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices	Pt 385, 500 Ω		Comparison to Multi-Product Calibrator
	(-200 to -80) $^{\circ}\text{C}$	0.11 $^{\circ}\text{C}$	
	(-80 to 0) $^{\circ}\text{C}$	0.12 $^{\circ}\text{C}$	
	(0 to 100) $^{\circ}\text{C}$	0.12 $^{\circ}\text{C}$	
	(100 to 260) $^{\circ}\text{C}$	0.12 $^{\circ}\text{C}$	
	(260 to 300) $^{\circ}\text{C}$	0.13 $^{\circ}\text{C}$	
	(300 to 400) $^{\circ}\text{C}$	0.13 $^{\circ}\text{C}$	
	(400 to 600) $^{\circ}\text{C}$	0.14 $^{\circ}\text{C}$	
	(600 to 630) $^{\circ}\text{C}$	0.15 $^{\circ}\text{C}$	
	Pt 385, 1 000 Ω		
	(-200 to -80) $^{\circ}\text{C}$	0.11 $^{\circ}\text{C}$	
	(-80 to 0) $^{\circ}\text{C}$	0.11 $^{\circ}\text{C}$	
	(0 to 100) $^{\circ}\text{C}$	0.11 $^{\circ}\text{C}$	
	(100 to 260) $^{\circ}\text{C}$	0.12 $^{\circ}\text{C}$	
	(260 to 300) $^{\circ}\text{C}$	0.12 $^{\circ}\text{C}$	
	(300 to 400) $^{\circ}\text{C}$	0.13 $^{\circ}\text{C}$	
	(400 to 600) $^{\circ}\text{C}$	0.13 $^{\circ}\text{C}$	
	(600 to 630) $^{\circ}\text{C}$	0.26 $^{\circ}\text{C}$	
	Pt 385, 120 Ω (Ni 120)		
	(-80 to 0) $^{\circ}\text{C}$	0.13 $^{\circ}\text{C}$	
	(0 to 100) $^{\circ}\text{C}$	0.13 $^{\circ}\text{C}$	
	(100 to 260) $^{\circ}\text{C}$	0.18 $^{\circ}\text{C}$	
	Cu 427, 10 Ω		
	(-100 to 260) $^{\circ}\text{C}$	0.32 $^{\circ}\text{C}$	
Electrical Simulation of RTD Indicating devices ¹	Pt Ni 120 (100 Ω)		Comparison to Precision Temperature Calibrator
	(-80 to 260) $^{\circ}\text{C}$	0.26 $^{\circ}\text{C}$	
	Pt 385 (100 Ω)		
	(-200 to 800) $^{\circ}\text{C}$	0.37 $^{\circ}\text{C}$	
	Pt 392 (100 Ω)		
	(-200 to 630) $^{\circ}\text{C}$	0.34 $^{\circ}\text{C}$	
	Pt JIS (100 Ω)		
	(-200 to 360) $^{\circ}\text{C}$	0.34 $^{\circ}\text{C}$	
	Pt 385 (200 Ω)		
	(-200 to 250) $^{\circ}\text{C}$	0.26 $^{\circ}\text{C}$	
	(250 to 630) $^{\circ}\text{C}$	0.82 $^{\circ}\text{C}$	
	Pt 385 (500 Ω)		
	(-200 to 500) $^{\circ}\text{C}$	0.34 $^{\circ}\text{C}$	
	(500 to 630) $^{\circ}\text{C}$	0.43 $^{\circ}\text{C}$	
	Pt 385 (1 000 Ω)		
	(-200 to 100) $^{\circ}\text{C}$	0.26 $^{\circ}\text{C}$	
	(100 to 630) $^{\circ}\text{C}$	0.26 $^{\circ}\text{C}$	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks ²	Up to 4 in (4 to 20) in	$(1.2 + 3.7L) \mu\text{in}$ $(0.27 + 1.8L) \mu\text{in}$	Mechanical Comparison using a Gage Block Comparator and Master Gage Blocks
Gage Block Comparator	Up to 20 in	1.1 μin	Comparison to Master Gage Blocks
Pin Gages ¹	Up to 2 in	28 μin	Non-contact Comparison using a Laser Micrometer and Master Plug Gages
Plain Plug Gages ^{1,2}	Up to 10 in (10 to 36) in	$(15 + 1.2L) \mu\text{in}$ $(18 + 1.3L) \mu\text{in}$	Comparison to Bench Micrometer, Gage Blocks
Plain Ring Gages ²	(0.040 to 12) in	$(15 + 1.8L) \mu\text{in}$	Comparison to I.D. Comparator, Gage Blocks
Thread Measuring Wires	Up to 0.5 in	17 μin	Comparison to Bench Micrometer, Gage Blocks, Master Plug Gage
Thread Plug Gages ¹ Major Diameter Pitch Diameter	Up to 10 in	64 μin	Comparison to Bench Micrometer, Thread Wires, Gage Blocks
Thread Plug Gages Tapered Pitch Diameter	Up to 10 in	72 μin	Comparison to Bench Micrometer, Thread Wires, Gage Blocks
Thread Ring Gages ¹	Up to 10 in	72 μin	Comparison to Thread Set Plugs
End Rod Standards ^{1,2}	Up to 60 in	$(14 + 2.4L) \mu\text{in}$	Comparison to Precision Measuring Machine, Gage Blocks
Calipers ^{1,2}	Up to 40 in 40 to 72 in	$(300 + 0.16L) \mu\text{in}$ $(290 + 0.4L) \mu\text{in}$	Comparisons to Caliper Master, Gage Blocks, Surface Plate
Caliper Master ^{1,2}	Up to 12 in (12 to 40) in	68.4 μin $(43 + 1.8L) \mu\text{in}$	Comparisons to Gage Blocks, Precision Indicator, Surface Plate

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Height Gages ^{1,2}	Up to 40 in	$(91 + 0.5L) \mu\text{in}$	Comparison to Gage Blocks, Precision Indicator, Surface Plate
Indicators ¹	Up to 6 in	$42 \mu\text{in}$	Comparison to Bench Micrometer, Gage Blocks
LVDTs ¹	Up to 0.05 in	$16 \mu\text{in}$	Comparison to Bench Micrometer, Gage Blocks
Depth Micrometers ¹	Up to 12 in	$56 \mu\text{in}$	Comparison to Gage Blocks, Surface Plate
Outside Micrometers ^{1,2}	Up to 24 in (24 to 78) in	$(27 + 1.8L) \mu\text{in}$ $(580 + 0.2L) \text{uin}$	Comparison to Gage Blocks, Optical Flat
Inside Micrometers ^{1,2}	Up to 40 in (Rod Length)	$(56 + 0.8L) \mu\text{in}$	Comparison to Precision Measuring Machine, Gage Blocks
Laser Micrometers ¹	Up to 2 in	$18 \mu\text{in}$	Comparison to Precision Master Plain Plugs
Supermicrometers / Bench Micrometers ^{1,3}	Up to 1 in Travel Up to 2 in Travel	$7.1 \mu\text{in}$ $7.5 \mu\text{in}$	Comparison to Gage Blocks
ID/OD Comparator ^{1,2} (Comparative)	Up to 12 in	$(3.5 + 1.3L) \mu\text{in}$	Comparison to Gage Blocks
Radius Gages	Up to 1 in	$220 \mu\text{in}$	Direct Measure using Precision Vision System
Scales / Rules / Tapes ²	Up to 12 ft	$(160 + 0.71L) \mu\text{in}$	Direct Measure using Precision Vision System

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Surface Plates ^{1,2}			
Overall Flatness	Up to (120 X 120) in	$(64 + 0.53DL) \mu\text{in}$	In accordance with ASME B89.3.7 using Laser Optics System
Local Area Flatness (Repeat Reading)	Up to (24 x 24) in Up to (120 x 120) in	42 μin 36 μin	Mahr Indicator Repeat-o-Meter, Precision Indicator
Optical Comparators ¹			
Angle	Up to 360°	0.04°	Comparisons to Precision Balls, Glass Scales, Gage Blocks
Magnification	(5 to 100) X	470 μin	
Radius	Up to 12 in	140 μin	
Squareness of Table	Up to 12 in	160 μin	
X-Y Axis Linearity	Up to 12 in	150 μin	
Vision Systems ¹			
Angle	Up to 12 in	0.021°	Comparison to Precision Balls, Glass Scales, Gage Blocks
Radius	Up to 12 in	61 μin	
Squareness of Table	Up to 12 in	39 μin	
X-Y Axis Linearity	Up to 12 in	63 μin	

Mass and Mass-Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Torque Tools ^{1,2}	Up to 50 ozf·in Up to 1 000 lbf·in Up to 1 000 lbf·ft Up to 2 000 lbf·ft	0.5 ozf·in $(0.2 + 0.004 \text{ 1X}) \text{ lbf·in}$ $(0.3 + 0.003 \text{ 1X}) \text{ lbf·ft}$ 22 lbf·ft	Comparison to Torque Tester
Torque Transducers ¹	(5 to 50) ozf·in (4 to 50) lbf·in (30 to 400) lbf·in (80 to 1 000) lbf·in (20 to 250) lbf·ft (60 to 600) lbf·ft (100 to 1 000) lbf·ft (200 to 2 000) lbf·ft	0.062 ozf·in 0.062 lbf·in 0.5 lbf·in 1.3 lbf·in 0.31 lbf·ft 0.74 lbf·ft 1.3 lbf·ft 2.5 lbf·ft	Comparison to Master Weights, Torque Arms

Mass and Mass-Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Torque Arms	Up to 2.5 in (2.5 to 5) in (5 to 10) in (10 to 20) in	59 μ in 88 μ in 160 μ in 300 μ in	Comparison to Gage Blocks, Master Plugs, Bench Micrometer
Durometer Calibrators (A, B, O, C, D, DO)	(0.55 to 45.45) N	0.016 N	Comparison to Force Gage, Load Cell
Durometers (Types A, B, C, D, DO, O) Spring Force Indenter Dimensions Shape Length	(0.55 to 8.05) N Up to 45.45 N (30 to 35)° (0.096 to 0.1) in	0.9 Duro Points 0.9 Duro Points 0.03° 66 μ in	Direct verification per internal cal procedure using Durometer Calibrator Vision system
Rockwell Hardness Testers ¹	HRBW Low Medium High HRC Low Medium High	0.8 HRBW 0.81 HRBW 0.85 HRBW 0.86 HRC 0.83 HRC 0.82 HRC	Indirect Verification per ASTM E18 using Hardness Test Blocks.
Rockwell Superficial Hardness Testers ¹	HR15N HR30N HR45N HR15TW HR30TW HR45TW	0.94 HR15N 0.91 HR30N 0.97 HR45N 0.93 HRTW 0.92 HRTW 0.94 HRTW	Indirect Verification per ASTM E18 using Hardness Test Blocks.
Microhardness Testers ¹	Repeatability under forces P (gf) 500 $\leq$ P $\leq$ 1 000 557 < HK > 685 Error 571 < HV > 790 Error	 1.3 μ m 0.08 μ m 2.9 μ m 0.9 μ m	Indirect Verification per ASTM E92 using Hardness Test Blocks.
Force Gages ^{1,2}	Up to 500 lbf Up to 1 000 lbf	(0.02 + 0.002X) lbf (0.12 + 0.004X) lbf	Comparison to Class F Weights, Force Tester

Mass and Mass-Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Cable Tensiometers ²	Up to 1 000 lbf	(0.34 + 0.007 5X) lbf	Comparison to Force Tester
Pressure Gages, Pressure Transducers ¹	Up to 5 psig Up to 200 psig Up to 1 000 psig Up to 3 000 psig Up to 7 000 psig Up to 10 000 psig	0.002 4 psi 0.14 psi 0.53 psi 0.67 psi 8.5 psi 16 psi	Comparison to Master Pressure Gages
Vacuum Devices	Up to 30 inHg	0.12 inHg	Comparison to Vacuum Test Gage
Scales and Balances ^{1,4}	Up to 120 g (120 to 220) g	93 µg 0.83 mg	ASTM E617 Class 0 Weights and internal calibration procedure utilized in the calibration of the weighing system.
Scales and Balances ^{1,4}	(220 to 510) g	1.2 mg	ASTM E617 Class 4 Weights and internal calibration procedure utilized in the calibration of the weighing system.
Scales and Balances ^{1,4}	(510 to 3 000) g	13 mg	ASTM E617 Class 1 Weights and internal calibration procedure utilized in the calibration of the weighing system.
Scales and Balances ^{1,2,4}	Up to 1 000 lb	(0.26 + 0.001 7X) lb	NIST Class F Weights and internal calibration procedure utilized in the calibration of the weighing system.

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Thermometers ¹	(35 to 100) °C (100 to 250) °C (250 to 375) °C	0.32 °C 0.49 °C 0.67 °C	Direct Measure using Thermal Dry Well
Temperature Thermometers ¹	(-200 to 1 200) °C	1.3 °C	Comparison to Precision Temperature Calibrator w/ Type J Thermocouple Probe

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Stopwatch /Timers	Up to 24 hr	490 ms	Comparison to N.I.S.T. Time Signal
Frequency – Source	Up to 100 Hz 100 Hz to 1 kHz (1 to 100) kHz (100 to 500) kHz	0.33 mHz 2.9 mHz 0.29 Hz 1.5 Hz	Comparison to Multi-Product Calibrator

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. DL = diagonal length in inches; L = length in inches; X = value of applied mass, force, or torque.
3. Uncertainty based on measurement over 1-inch lead screw.
4. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration. Calibration weighing system calibrated in accordance with internal calibration procedure CAL-SB290.
5. Unless otherwise specified in the far-right hand column, the calibration procedure being utilized by the laboratory was written internally.
6. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1160.



Jason Stine, Vice President